



Republic of the Philippines
Province of South Cotabato
City of Koronadal

Bids and Awards Committee

E-mail Address: lgubackor@gmail.com Website: www.philgeps.gov.ph and www.koronadal.gov.ph

SUPPLEMENTAL BID BULLETIN NO. 2

23 MAY 2024

PROCUREMENT FOR THE SUPPLY & DELIVERY OF ONE (1) UNIT AMBULANCE W/ EQUIPMENTS & SUPPLIES, CITY OF KORONADAL

This Bid Bulletin is being issued to announce the amendment's made in **ITB-KOR: 24-05-0058 CHO - PROCUREMENT FOR THE SUPPLY & DELIVERY OF ONE (1) UNIT AMBULANCE W/ EQUIPMENTS & SUPPLIES, CITY OF KORONADAL** with PhilGEPS Reference No. 10825533. This Bid Bulletin shall form an integral part of the bidding documents. The following are the changes in the bidding documents:

Amendments made is in reference to the PRE-BIDDING CONFERENCE conducted last 14 MAY 2023.

| REFERENCE | AMMENDMENTS/INCLUSIONS |
|--|--|
| Section VII. TECHNICAL SPECIFICATIONS Page 24, Specification AMBULANCE EQUIPMENT & ACCESSORIES | Section VII. TECHNICAL SPECIFICATIONS Page 24, Specification AMBULANCE EQUIPMENT & ACCESSORIES (ADDITIONAL SPECIFICATIONS FOR <u>ITEM 15 – 1 UNIT PATIENT STRETCHER (FOLDABLE)</u> 15.A – MATERIAL: ALUMINUM ALLOY FRAME WITH STAINLESS COMPONENTS FOR DURABILITY, LIGHTWEIGHT CONSTRUCTION, AND RESISTANCE TO CORROSION. WITH MATTRESS. |
| | Section VII. TECHNICAL SPECIFICATIONS Page 25, Specification AMBULANCE EQUIPMENT & ACCESSORIES (ADDITIONAL SPECIFICATIONS FOR <u>ITEM 15 – 1 UNIT PATIENT STRETCHER (FOLDABLE)</u> 15.B – DIMENSIONS: L: APPROXIMATELY 72 INCHES (183 cm) W: APPROXIMATELY 24 INCHES (61 cm) H: ADJUSTABLE, TYPICALLY RANGING FROM 20 to 34 INCHES (51 to 86 cm) |
| | Section VII. TECHNICAL SPECIFICATIONS Page 25, Specification AMBULANCE EQUIPMENT & ACCESSORIES (ADDITIONAL SPECIFICATIONS FOR <u>ITEM 15 – 1 UNIT PATIENT STRETCHER (FOLDABLE)</u> 15.C. – WEIGHT: MINIMUM – 227 kg; MAXIMUM – 318 kg |



Republic of the Philippines
Province of South Cotabato
City of Koronadal

Bids and Awards Committee

E-mail Address: lgubackor@gmail.com Website: www.philgeps.gov.ph and www.koronadal.gov.ph

| | |
|--|--|
| | <p>Section VII. TECHNICAL SPECIFICATIONS Page 25, Specification AMBULANCE EQUIPMENT & ACCESSORIES</p> <p><u>(ADDITIONAL SPECIFICATIONS FOR ITEM 15 – 1 UNIT PATIENT STRETCHER (FOLDABLE))</u></p> <p>15.D – ADJUSTABILITY: - BACKREST: ADJUSTABLE FROM FLAT TO AT LEAST 45 DEGREES. HEIGHT: ADJUSTABLE TO FACILITATE LOADING & UNLOADING. LEG SECTION: ADJUSTABLE</p> |
| | <p>Section VII. TECHNICAL SPECIFICATIONS Page 25, Specification AMBULANCE EQUIPMENT & ACCESSORIES</p> <p><u>(ADDITIONAL SPECIFICATIONS FOR ITEM 15 – 1 UNIT PATIENT STRETCHER (FOLDABLE))</u></p> <p>15.E – SAFETY FEATURES: - SIDE RAILS: REMOVABLE OR COLLAPSIBLE SIDE RAILS. RESTRAINTS: INTEGRATED STRAPS OR BELTS TO SECURE THE PATIENT SAFELY TO THE STRTCHER. BRAKES: WHEEL LOCKS REFLECTIVE MARKINGS: ENHANCE VISIBILITY, ESPECIALLY DURING LOW-LIGHT CONDITIONS OR EMERGENCIES.</p> |
| | <p>Section VII. TECHNICAL SPECIFICATIONS Page 25, Specification AMBULANCE EQUIPMENT & ACCESSORIES</p> <p><u>(ADDITIONAL SPECIFICATIONS FOR ITEM 15 – 1 UNIT PATIENT STRETCHER (FOLDABLE))</u></p> <p>15.F – COMPATABILITY: DESIGNED TO BE COMPATIBLE WITH STANDARD AMBULANCE MOUNTUNG SYSTEMS, ENSURING SECURE ATTACHMENT AND STABILITY DURING TRANSPORTATION.</p> |
| | <p>Section VII. TECHNICAL SPECIFICATIONS Page 25, Specification AMBULANCE EQUIPMENT & ACCESSORIES</p> <p><u>(ADDITIONAL SPECIFICATIONS FOR ITEM 15 – 1 UNIT PATIENT STRETCHER (FOLDABLE))</u></p> <p>15.G – MOBILITY: FOUR SWIVEL WHEELS WITH AT LEAST TWO (2) LOCKABLE SHEELS FOR STABILITY DURING LOADING & UNLOADING. – LARGE DIAMETER WHEELS</p> |



Republic of the Philippines
Province of South Cotabato
City of Koronadal

Bids and Awards Committee

E-mail Address: lgubackor@gmail.com Website: www.philgeps.gov.ph and www.koronadal.gov.ph

| | |
|--|--|
| | <p>Section VII. TECHNICAL SPECIFICATIONS Page 25, Specification AMBULANCE EQUIPMENT & ACCESSORIES</p> <p><u>(ADDITIONAL SPECIFICATIONS FOR ITEM 15 – 1 UNIT PATIENT STRETCHER (FOLDABLE))</u></p> <p>15.H – ACCESSIBILITY: - FOLDABLE OR REMOVABLE LEGS FOR EASY MANUEVERABILITY THROUGH NARROW SPACES - LOW-PROFILE DESIGN FOR SEAMLESS TRANSFER OF PATIENTS FROM STRETCHER TO AMBULANCE & VICE VERSA</p> |
| | <p>Section VII. TECHNICAL SPECIFICATIONS Page 25, Specification AMBULANCE EQUIPMENT & ACCESSORIES</p> <p><u>(ADDITIONAL SPECIFICATIONS FOR ITEM 15 – 1 UNIT PATIENT STRETCHER (FOLDABLE))</u></p> <p>15.I – ADDITIONAL FEATURES: - IV POLE MOUNTS FOR ADMINISTERING FLUIDS DURING TRANSPORT. - OXYGEN CYLINDER HOLDER FOR SECURING OXYGEN SUPPLY</p> |
| | <p>Section VII. TECHNICAL SPECIFICATIONS Page 25, Specification AMBULANCE EQUIPMENT & ACCESSORIES</p> <p><u>(ADDITIONAL SPECIFICATIONS FOR ITEM 16 – 1 UNIT SPINE BOARD)</u></p> <p><u>16.A – MATERIAL: LIGHTWEIGHT; DURABLE LIKE HIGH-DENSITY POLYTHYLENE (HDPE)</u></p> |
| | <p>Section VII. TECHNICAL SPECIFICATIONS Page 25, Specification AMBULANCE EQUIPMENT & ACCESSORIES</p> <p><u>(ADDITIONAL SPECIFICATIONS FOR ITEM 16 – 1 UNIT SPINE BOARD)</u></p> <p>16.B – DIMENSIONS: L:183cm W:41 to 46cm</p> |
| | <p>Section VII. TECHNICAL SPECIFICATIONS Page 25, Specification AMBULANCE EQUIPMENT & ACCESSORIES</p> <p><u>(ADDITIONAL SPECIFICATIONS FOR ITEM 16 – 1 UNIT SPINE BOARD)</u></p> |



Republic of the Philippines
Province of South Cotabato
City of Koronadal

Bids and Awards Committee

E-mail Address: lgubackor@gmail.com Website: www.philgeps.gov.ph and www.koronadal.gov.ph

| | |
|--|--|
| | <p>16.C – WEIGHT CAPACITY: MINIMUM: 500 POUNDS (227 KG)</p> |
| | <p>Section VII. TECHNICAL SPECIFICATIONS Page 25, Specification AMBULANCE EQUIPMENT & ACCESSORIES</p> <p>(ADDITIONAL SPECIFICATIONS FOR <u>ITEM 16 – 1 UNIT SPINE BOARD</u>)</p> <p>16.D – SAFETY FEATURES: - RADIOLUCENT: TRANSPARENT TO X-RAYS FOR MEDICAL IMAGING WITHOUT THE NEED TO REMOVE THE PATIENT FROM THE BOARD. - IMMOBILIZATION: FEATURES STRAPS OR BELTS TO SECURE ATTACHMENT & STABILITY DURING TRANSPORT</p> |
| | <p>Section VII. TECHNICAL SPECIFICATIONS Page 25, Specification AMBULANCE EQUIPMENT & ACCESSORIES</p> <p>(ADDITIONAL SPECIFICATIONS FOR <u>ITEM 16 – 1 UNIT SPINE BOARD</u>)</p> <p>16.E – COMPATIBILITY: DESIGNED TO FIT STANDARD AMBULANCE STRETCHER MOUNTING SYSTEMS FOR SECURE ATTACHMENT & STABILITY DURING TRANSPORT.</p> |
| | <p>Section VII. TECHNICAL SPECIFICATIONS Page 25, Specification AMBULANCE EQUIPMENT & ACCESSORIES</p> <p>(ADDITIONAL SPECIFICATIONS FOR <u>ITEM 16 – 1 UNIT SPINE BOARD</u>)</p> <p>16.F – MOBILITY: LIGHTWEIGHT DESIGN FOR EASE OF HANDLING & MANUEVERABILITY DURING PATIENT TRANSPORT.</p> |
| | <p>Section VII. TECHNICAL SPECIFICATIONS Page 25, Specification AMBULANCE EQUIPMENT & ACCESSORIES</p> <p>(ADDITIONAL SPECIFICATIONS FOR <u>ITEM 16 – 1 UNIT SPINE BOARD</u>)</p> <p>16.G – ACCESSIBILITY: LOW-PROFILE DESIGN FOR EASY INSERTION & REMOVAL FROM AMBULANCES</p> |



Republic of the Philippines
Province of South Cotabato
City of Koronadal

Bids and Awards Committee

E-mail Address: lgubackor@gmail.com Website: www.philgeps.gov.ph and www.koronadal.gov.ph

| | |
|--|---|
| | <p>Section VII. TECHNICAL SPECIFICATIONS Page 25, Specification AMBULANCE EQUIPMENT & ACCESSORIES</p> <p>(ADDITIONAL SPECIFICATIONS FOR <u>ITEM 17</u> <u>- 1 UNIT PATIENT MONITOR</u></p> <p><u>17.A – TYPE: PORTABLE MULTI-PARAMETER PATIENT MONITOR CAPABLE OF MONITORING VITAL SIGNS IN REAL-TIME DURING AMBULANCE TRANSPORT</u></p> |
| | <p>Section VII. TECHNICAL SPECIFICATIONS Page 25, Specification AMBULANCE EQUIPMENT & ACCESSORIES</p> <p>(ADDITIONAL SPECIFICATIONS FOR <u>ITEM 17</u> <u>- 1 UNIT PATIENT MONITOR</u></p> <p>17.B – PARAMETERS: TYPICALLY MEASURES VITAL SIGNS SUCH AS: - ELECTROCARDIOGRAM (ECG) FOR HEART RHYTHM & FUNCTION. - BLOOD PRESSURE (NON-INVASIVE OR INVASIVE) FOR MONITORING CIRCULATORY STATUS. - PULSE OXIMETRY (SpO2) FOR BLOOD OXYGEN SATURATION LEVELS. - RESPIRATORY RATE FOR MONITORING BREATHING. - TEMPERATURE FOR ASSESSING PATIENT’S BODY TEMPERATURE.</p> |
| | <p>Section VII. TECHNICAL SPECIFICATIONS Page 25, Specification AMBULANCE EQUIPMENT & ACCESSORIES</p> <p>(ADDITIONAL SPECIFICATIONS FOR <u>ITEM 17</u> <u>- 1 UNIT PATIENT MONITOR</u></p> <p>17.C – DISPLAY: HIGH-RESOLUTION COLOR DISPLAY SCREEN; 8-10 INCHES</p> |
| | <p>Section VII. TECHNICAL SPECIFICATIONS Page 25, Specification AMBULANCE EQUIPMENT & ACCESSORIES</p> <p>(ADDITIONAL SPECIFICATIONS FOR <u>ITEM 17</u> <u>- 1 UNIT PATIENT MONITOR</u></p> <p>17.D – ALARMS: AUDIBLE & VISUAL ALARMS FOR ABNORMAL VITAL SIGNS, INCLUDING CUSTOMIZABLE ALARM LIMITS BASED ON</p> |



Republic of the Philippines
Province of South Cotabato
City of Koronadal

Bids and Awards Committee

E-mail Address: lgubackor@gmail.com Website: www.philgeps.gov.ph and www.koronadal.gov.ph

| | |
|--|---|
| | <p>PATIENT CONDITION & MEDICAL PROTOCOLS</p> |
| | <p>Section VII. TECHNICAL SPECIFICATIONS Page 25, Specification AMBULANCE EQUIPMENT & ACCESSORIES</p> <p>(ADDITIONAL SPECIFICATIONS FOR <u>ITEM 17 – 1 UNIT PATIENT MONITOR</u>)</p> <p>17.E – BATTERY LIFE: LONG BATTERY LIFE TO ENSURE CONTINUOUS MONITORING DURING AMBULANCE TRANSPORT, TYPICALLY LASTING FOR SEVERAL HOURS ON A SINGLE CHARGE.</p> |
| | <p>Section VII. TECHNICAL SPECIFICATIONS Page 25, Specification AMBULANCE EQUIPMENT & ACCESSORIES</p> <p>(ADDITIONAL SPECIFICATIONS FOR <u>ITEM 17 – 1 UNIT PATIENT MONITOR</u>)</p> <p>17.F – CONNECTIVITY: WIRED AND/OR WIRELESS CONNECTIVITY OPTIONS FOR DATA TRANSMISSION & INTEGRATION W/ ELECTRONIC MEDICAL RECORDS (EMR) SYSTEMS. - COMPATABILITY W/ TELEMETRY SYSTEMS FOR REMOTE MONITORING & COMMUNICATION W/ HOSPITAL STAFF</p> |
| | <p>Section VII. TECHNICAL SPECIFICATIONS Page 25, Specification AMBULANCE EQUIPMENT & ACCESSORIES</p> <p>(ADDITIONAL SPECIFICATIONS FOR <u>ITEM 17 – 1 UNIT PATIENT MONITOR</u>)</p> <p>17.G – DURABILITY: SHOCK-RESISTANT & RUGGED CONSTRUCTION</p> |
| | <p>Section VII. TECHNICAL SPECIFICATIONS Page 25, Specification AMBULANCE EQUIPMENT & ACCESSORIES</p> <p>(ADDITIONAL SPECIFICATIONS FOR <u>ITEM 17 – 1 UNIT PATIENT MONITOR</u>)</p> <p>17.H – USER INTERFACE: INTUATIVE USER INTERFACE W/ TOUGHSCREEN OR TACTILE BUTTONS FOR EASY</p> |



Republic of the Philippines
Province of South Cotabato
City of Koronadal

Bids and Awards Committee

E-mail Address: lgubackor@gmail.com Website: www.philgeps.gov.ph and www.koronadal.gov.ph

| | |
|--|--|
| | <p>Section VII. TECHNICAL SPECIFICATIONS Page 25, Specification AMBULANCE EQUIPMENT & ACCESSORIES</p> <p>(ADDITIONAL SPECIFICATIONS FOR <u>ITEM 18</u> <u>- 1 UNIT SUCTION MACHINE</u>)</p> <p><u>18.A – TYPE: PORTABLE ELECTRIC SUCTION</u></p> |
| | <p>Section VII. TECHNICAL SPECIFICATIONS Page 25, Specification AMBULANCE EQUIPMENT & ACCESSORIES</p> <p>(ADDITIONAL SPECIFICATIONS FOR <u>ITEM 18</u> <u>- 1 UNIT SUCTION MACHINE</u>)</p> <p>18.B – SUCTION POWER: TYPICALLY PROVIDES ADJUSTABLE SUCTION PRESSURE RANGING FROM 0 TO 600+ mmHg</p> |
| | <p>Section VII. TECHNICAL SPECIFICATIONS Page 25, Specification AMBULANCE EQUIPMENT & ACCESSORIES</p> <p>(ADDITIONAL SPECIFICATIONS FOR <u>ITEM 18</u> <u>- 1 UNIT SUCTION MACHINE</u>)</p> <p>18.C – FLOW RATE: 20 TO 40 LITERS PER MINUTE (LPM)</p> |
| | <p>Section VII. TECHNICAL SPECIFICATIONS Page 25, Specification AMBULANCE EQUIPMENT & ACCESSORIES</p> <p>(ADDITIONAL SPECIFICATIONS FOR <u>ITEM 18</u> <u>- 1 UNIT SUCTION MACHINE</u>)</p> <p>18.D – POWER SOURCE: INTERNAL RECHARGEABLE BATTERY AND/OR DIRECT CONNECTION TO THE AMBULANCE'S ELECTRICAL SYSTEM FOR CONTINUOUS OPERATION DURING TRANSPORT.</p> |
| | <p>Section VII. TECHNICAL SPECIFICATIONS Page 25, Specification AMBULANCE EQUIPMENT & ACCESSORIES</p> <p>(ADDITIONAL SPECIFICATIONS FOR <u>ITEM 18</u> <u>- 1 UNIT SUCTION MACHINE</u>)</p> <p>18.E – VACUUM CONTROL: ADJUSTABLE VACUUM CONTROL SETTINGS</p> |



Republic of the Philippines
Province of South Cotabato
City of Koronadal

Bids and Awards Committee

E-mail Address: lgubackor@gmail.com Website: www.philgeps.gov.ph and www.koronadal.gov.ph

| | |
|--|--|
| | <p>Section VII. TECHNICAL SPECIFICATIONS Page 25, Specification AMBULANCE EQUIPMENT & ACCESSORIES</p> <p><u>(ADDITIONAL SPECIFICATIONS FOR ITEM 18 – 1 UNIT SUCTION MACHINE)</u></p> <p>18.F – COLLECTION CANISTER: REMOVABLE & AUTOCLAVABLE COLLECTION CANISTER W/ A CAPACITY OF AT LEAST 1,000 MILLILITERS (ml)</p> |
| | <p>Section VII. TECHNICAL SPECIFICATIONS Page 25, Specification AMBULANCE EQUIPMENT & ACCESSORIES</p> <p><u>(ADDITIONAL SPECIFICATIONS FOR ITEM 18 – 1 UNIT SUCTION MACHINE)</u></p> <p>18.G – FILTER SYSTEM: INTEGRATED BACTERIAL FILTER SYSTEM</p> |
| | <p>Section VII. TECHNICAL SPECIFICATIONS Page 25, Specification AMBULANCE EQUIPMENT & ACCESSORIES</p> <p><u>(ADDITIONAL SPECIFICATIONS FOR ITEM 18 – 1 UNIT SUCTION MACHINE)</u></p> <p>18.H – DURABILITY: DURABLE CONSTRUCTION W/ SHOCK RESISTANT</p> |
| | <p>Section VII. TECHNICAL SPECIFICATIONS Page 25, Specification AMBULANCE EQUIPMENT & ACCESSORIES</p> <p><u>(ADDITIONAL SPECIFICATIONS FOR ITEM 18 – 1 UNIT SUCTION MACHINE)</u></p> <p>18.I – USER INTERFACE: INTUITIVE CONTROL PANEL W/ EASY-TO-USE SWITCHES OR BUTTONS FOR ADJUSTING SUCTION SETTINGS & MONITORING OPERATION.</p> |
| | <p>Section VII. TECHNICAL SPECIFICATIONS Page 25, Specification AMBULANCE EQUIPMENT & ACCESSORIES</p> <p><u>(ADDITIONAL SPECIFICATIONS FOR ITEM 18 – 1 UNIT SUCTION MACHINE)</u></p> <p>18.J – ACCESSORIES: SUCTION CATHETERS, TUBING & CONNECTORS FOR VERSATILE SUCTIONING</p> |



Republic of the Philippines
Province of South Cotabato
City of Koronadal

Bids and Awards Committee

E-mail Address: lgubackor@gmail.com Website: www.philgeps.gov.ph and www.koronadal.gov.ph

| | |
|--|---|
| | <p>Section VII. TECHNICAL SPECIFICATIONS Page 25, Specification AMBULANCE EQUIPMENT & ACCESSORIES</p> <p><u>(ADDITIONAL SPECIFICATIONS FOR ITEM 18 – 1 UNIT SUCTION MACHINE)</u></p> <p>18.K – SAFETY FEATURES: OVERPRESSURE PROTECTION & MINIMIZE THE RISK OF TISSUE DAMAGE. AUDIBLE & VISUAL ALARMS TO INDICATE LOW BATTERY, CANISTER FULL, OR SYSTEM MALFUNCTIONS.</p> |
| | <p>Section VII. TECHNICAL SPECIFICATIONS Page 25, Specification AMBULANCE EQUIPMENT & ACCESSORIES</p> <p><u>(ADDITIONAL SPECIFICATIONS FOR ITEM 19 – 1 UNIT AUTOMATIC EXTERNAL DEFIBRILLATOR (AED))</u></p> <p>19.A – TYPE: AUTOMATED EXTERNAL DEFIBRILLATOR (AED)</p> |
| | <p>Section VII. TECHNICAL SPECIFICATIONS Page 25, Specification AMBULANCE EQUIPMENT & ACCESSORIES</p> <p><u>(ADDITIONAL SPECIFICATIONS FOR ITEM 19 – 1 UNIT AUTOMATIC EXTERNAL DEFIBRILLATOR (AED))</u></p> <p>19.B – OPERATION: FULLY AUTOMATED OPERATION W/ CLEAR AUDIO & VISUAL PROMPTS</p> |
| | <p>Section VII. TECHNICAL SPECIFICATIONS Page 25, Specification AMBULANCE EQUIPMENT & ACCESSORIES</p> <p><u>(ADDITIONAL SPECIFICATIONS FOR ITEM 19 – 1 UNIT AUTOMATIC EXTERNAL DEFIBRILLATOR (AED))</u></p> <p>19.C – WAVE FORM: BIPHASIC WAVEFORM TECHNOLOGY</p> |
| | <p>Section VII. TECHNICAL SPECIFICATIONS Page 25, Specification AMBULANCE EQUIPMENT & ACCESSORIES</p> |



Republic of the Philippines
 Province of South Cotabato
 City of Koronadal

Bids and Awards Committee

E-mail Address: lgubackor@gmail.com Website: www.philgeps.gov.ph and www.koronadal.gov.ph

| | |
|--|---|
| | <p>(ADDITIONAL SPECIFICATIONS FOR <u>ITEM 19 – 1 UNIT AUTOMATIC EXTERNAL DEFIBRILLATOR (AED)</u>)</p> <p>19.D – ENERGY LEVELS: ADJUSTABLE ENERGY LEVELS FOR DEFIBRILLATION SHOCKS, TYPICALLY RANGING FROM 100 TO 360 JOULES</p> |
| | <p>Section VII. TECHNICAL SPECIFICATIONS Page 25, Specification AMBULANCE EQUIPMENT & ACCESSORIES</p> <p>(ADDITIONAL SPECIFICATIONS FOR <u>ITEM 19 – 1 UNIT AUTOMATIC EXTERNAL DEFIBRILLATOR (AED)</u>)</p> <p>19.E – SEVSITIVITY: ADVANCED RHYTHM ANALYSIS ALGORITMS: VENTRICULAR FIBRILLATION (VF) & PULSELESS VENTRICULAR TACHYCARDIA (VT)</p> |
| | <p>Section VII. TECHNICAL SPECIFICATIONS Page 25, Specification AMBULANCE EQUIPMENT & ACCESSORIES</p> <p>(ADDITIONAL SPECIFICATIONS FOR <u>ITEM 19 – 1 UNIT AUTOMATIC EXTERNAL DEFIBRILLATOR (AED)</u>)</p> <p>19.F – ELECTRODE PADS: PRE-CONNECTED OR EASILY ATTACHABLE ELECTRODE PADS W/ CLEAR PLACEMENTS DIAGRAMS</p> |
| | <p>Section VII. TECHNICAL SPECIFICATIONS Page 25, Specification AMBULANCE EQUIPMENT & ACCESSORIES</p> <p>(ADDITIONAL SPECIFICATIONS FOR <u>ITEM 19 – 1 UNIT AUTOMATIC EXTERNAL DEFIBRILLATOR (AED)</u>)</p> <p>19.G – BATTERY LIFE: LONG-LASTING BATTERY LIFE TO SUPPORT MULTIPLE RESCUE ATTEMPTS, TYPICALLY CAPABLE OF DELIVERING A MINIMUM OF 100 SHOCKS OR OPERATING CONTINUOUSLY FOR AT LEASE 8 HOURS</p> |
| | <p>Section VII. TECHNICAL SPECIFICATIONS Page 25, Specification AMBULANCE EQUIPMENT & ACCESSORIES</p> |



Republic of the Philippines
 Province of South Cotabato
 City of Koronadal

Bids and Awards Committee

E-mail Address: lgubackor@gmail.com Website: www.philgeps.gov.ph and www.koronadal.gov.ph

| | |
|--|--|
| | <p>(ADDITIONAL SPECIFICATIONS FOR <u>ITEM 19 – 1 UNIT AUTOMATIC EXTERNAL DEFIBRILLATOR (AED)</u>)</p> <p>19.H – USER INTERFACE: INTUATIVE USER INTERFACE W/ A RESPONSIVE TOUCHSCREEN OR KEYPAD FOR EASY OPERATION</p> |
| | <p>Section VII. TECHNICAL SPECIFICATIONS Page 25, Specification AMBULANCE EQUIPMENT & ACCESSORIES</p> <p>(ADDITIONAL SPECIFICATIONS FOR <u>ITEM 19 – 1 UNIT AUTOMATIC EXTERNAL DEFIBRILLATOR (AED)</u>)</p> <p>19.I – ACCESSORIES: DISPOSABLE ELECTRODE PADS, BATTERIES, CARRYING CASES & INSTRUCTION MANUALS FOR COMPREHENSIVE AED DEPLOYMENT</p> |
| | <p>Section VII. TECHNICAL SPECIFICATIONS Page 25, Specification AMBULANCE EQUIPMENT & ACCESSORIES</p> <p>(ADDITIONAL SPECIFICATIONS FOR <u>ITEM 20 – 1 UNIT OXYGEN TANK 20lbs W/ REGULATOR</u>)</p> <p>20.A – DIMENSIONS: HEIGHT: 90cm x CIRCUMFERENCE: 50cm</p> |
| | <p>Section VII. TECHNICAL SPECIFICATIONS Page 25, Specification AMBULANCE EQUIPMENT & ACCESSORIES</p> <p>(ADDITIONAL SPECIFICATIONS FOR <u>ITEM 21 – 1 UNIT DASHBOARD CAMERA</u>)</p> <p>-FRONT-FACING CAMERA: FIELD OF VIEW: 121° (SEMI-WIDE ANGLE) FRONT-FACING CAMERA RESOLUTION: HIGH DEFINITION (1080p) FRAME RATE: 30 fps LOW LIGHT: HDR TO OPTIMIZE FRO BRIGHT & LOW LIGHT CONDITIONS;</p> <p>AUDIO SPEAKER: BUILT-IN SPEAKER: INWAR FOR OPTIONAL VOICE & TONE BASED FEEDBACK</p> <p>STORAGE, CONNECTIVITY & COMPUTER: STORAGE: 24 HOURS OF 1080p FOOTAGE</p> |



Republic of the Philippines
Province of South Cotabato
City of Koronadal

Bids and Awards Committee

E-mail Address: lgubackor@gmail.com Website: www.philgeps.gov.ph and www.koronadal.gov.ph

| | |
|--|---|
| | <p>9CAN BE EXTENDED BY CONFIGURING VIDEO QUALITY)</p> <p>CONNECTIVITY: HIGH-SPEED 4G-LTE WIRELESS CONNECTIVITY. DATA SECURED VIA SSL 9256-BIT, MILITARY GRADE ENCRYPTION)</p> <p>EDGE-PROCESSING: HIGH-CAPACITY EDGE PROCESSOR TO ENABLE CONCURRENT NETWORKS FOR REAL-TIME ARTIFICIAL INTELLIGENCE</p> <p>AUTOMATIC UPLOAD: AUTOMATICALLY UPLOADS VIDEO FOOTAGE OF HARSH BRAKING, ACCELERATION, TURNING & CRASHES</p> <p>MOUNTING, ENCLOSURE & POWER: DIMENSIONS: ENCLOSURE: 108x41x61mm (4.2 x1.6 x2.4 in) CABLE: 4m (13ft) MOUNTING: MOUNTS TO WINDSHIELD VIA INDUSTRIAL GRADE ADHESIVE TAPE (INCLUDED) BATTERY: INTERNAL 1300mAh BATTERY FOR FALLBACK POWER WHEN VEHICLE IS OFF. SOFTWARE FEATURES: HD VIDEO PLAYBACK FROM ANY INTERNET-CONNECTED DEVICE SAFETY INCIDENT REPORTS, INCLUDING ACCIDENT RECONSTRUCTION DRIVER SAFETY RANKINGS W/ SCORECARDS ON-DEMAND RETRIEVAL (OVER THE AIR) BASED ON TIME OR LOCATION.</p> |
| | <p>Section VII. TECHNICAL SPECIFICATIONS Page 25, Specification AMBULANCE EQUIPMENT & ACCESSORIES</p> <p>(ADDITIONAL SPECIFICATIONS FOR <u>ITEM 22 – MEDICAL CABINET</u></p> <p><u>22.A SIZE & DIMENSIONS: 24-36 INCHES IN WIDTH, 12-18 INCHES IN DEPTH, & 24-36 INCHES IN HEIGHT</u></p> |
| | <p>Section VII. TECHNICAL SPECIFICATIONS Page 25, Specification AMBULANCE EQUIPMENT & ACCESSORIES</p> <p>(ADDITIONAL SPECIFICATIONS FOR <u>ITEM 22 – MEDICAL CABINET</u></p> |



Republic of the Philippines
 Province of South Cotabato
 City of Koronadal

Bids and Awards Committee

E-mail Address: lgubackor@gmail.com Website: www.philgeps.gov.ph and www.koronadal.gov.ph

| | |
|--|---|
| | <p>22.B – CONSTRUCTION MATERIAL: STAINLESS STEEL OR ALUMINUM TO WITHSTAND THE RIGORS OF TRANSPORTATION</p> |
| | <p>Section VII. TECHNICAL SPECIFICATIONS Page 25, Specification AMBULANCE EQUIPMENT & ACCESSORIES</p> <p>(ADDITIONAL SPECIFICATIONS FOR <u>ITEM 22 – MEDICAL CABINET</u>)</p> <p>22.C – STORAGE COMPARTMENTS: MULTIPLE COMPARTMENTS: ADJUSTABLE SHELVES</p> |
| | <p>Section VII. TECHNICAL SPECIFICATIONS Page 25, Specification AMBULANCE EQUIPMENT & ACCESSORIES</p> <p>(ADDITIONAL SPECIFICATIONS FOR <u>ITEM 22 – MEDICAL CABINET</u>)</p> <p>22.D – LOCKING MECHANISM: A KEYED LOCK OR COMBINATION LOCK</p> |
| | <p>Section VII. TECHNICAL SPECIFICATIONS Page 25, Specification AMBULANCE EQUIPMENT & ACCESSORIES</p> <p>(ADDITIONAL SPECIFICATIONS FOR <u>ITEM 22 – MEDICAL CABINET</u>)</p> <p>22.E VENTILATION: VENTILATION HOLES</p> |
| | <p>Section VII. TECHNICAL SPECIFICATIONS Page 25, Specification AMBULANCE EQUIPMENT & ACCESSORIES</p> <p>(ADDITIONAL SPECIFICATIONS FOR <u>ITEM 23 – DRESSING TRAY (ALUMINUM)</u>)</p> <p>23.A SIZE & DIMENSIONS: RANGE FROM AROUND 12 TO 18 INCHES IN LENGTH, 8 TO 12 INCHES IN WIDTH & 2 TO 4 INCHES IN HEIGHT.</p> |
| | <p>Section VII. TECHNICAL SPECIFICATIONS Page 25, Specification AMBULANCE EQUIPMENT & ACCESSORIES</p> <p>(ADDITIONAL SPECIFICATIONS FOR <u>ITEM 24 – SURGICAL SET</u>)</p> <p>24.A STERILE INSTRUMENTS: -SCALPEL HANDLE(S) W/ BLADES -SURGICAL SCISSORS 9E.G DISSECTING SCISSORS, BANDAGE SCISSORS)</p> |



Republic of the Philippines
Province of South Cotabato
City of Koronadal

Bids and Awards Committee

E-mail Address: lgubackor@gmail.com Website: www.philgeps.gov.ph and www.koronadal.gov.ph

| | |
|--|--|
| | <ul style="list-style-type: none">-FORCEPS (E.G TISSUE FORCEPS, DRESSING FORCEPS)-HEMOSTATS OR ARTERY FORCEPS-NEEDLE HOLDERS & SUTURE SCISSORSRETRACTORS (E.G SELF-RETAINING OR HAND-HELD)-SURGICAL PROBES & DILATORS-SURGICAL HOOKS OR ELEAVATORS |
| | <p>Section VII. TECHNICAL SPECIFICATIONS Page 25, Specification AMBULANCE EQUIPMENT & ACCESSORIES</p> <p>(ADDITIONAL SPECIFICATIONS FOR <u>ITEM 24 – SURGICAL SET</u>)</p> <p>24.B – STERILE DRESSINGS & WOUND MANAGEMENT SUPPLIES:</p> <ul style="list-style-type: none">-STERILE GAUZE PADS IN VARIOUS SIZES-ADHESIVE BANDAGES & TAPES-STERILE WOUND DRESSINGS (E.G NON-ADHERENT DRESSINGS, TRANSPARENT FILM DRESSINGS)-STERILE COTTON BALLS OR SWABS-STERILE SALINE SOLUTION FOR WOUND IRRIGATION-STERILE IRRIGATION SYRINGE OR BULB |
| | <p>Section VII. TECHNICAL SPECIFICATIONS Page 25, Specification AMBULANCE EQUIPMENT & ACCESSORIES</p> <p>(ADDITIONAL SPECIFICATIONS FOR <u>ITEM 24 – SURGICAL SET</u>)</p> <p>24.C – ANESTHESIA SUPPLIES:</p> <ul style="list-style-type: none">-LOCAL ANESTHETICS (E.G LIDOCAINE)-SYRINGE & NEEDLES FOR LOCAL ANRSTHESIA ADMINISTRATION |
| | <p>Section VII. TECHNICAL SPECIFICATIONS Page 25, Specification AMBULANCE EQUIPMENT & ACCESSORIES</p> <p>(ADDITIONAL SPECIFICATIONS FOR <u>ITEM 24 – SURGICAL SET</u>)</p> <p>24.D – STERILE DRAPES & COVERS:</p> <ul style="list-style-type: none">-STERILE TOWELS OR SHEETS FOR PATIENT DRAPING |



Republic of the Philippines
Province of South Cotabato
City of Koronadal

Bids and Awards Committee

E-mail Address: lgubackor@gmail.com Website: www.philgeps.gov.ph and www.koronadal.gov.ph

| | |
|--|--|
| | <p>Section VII. TECHNICAL SPECIFICATIONS Page 25, Specification AMBULANCE EQUIPMENT & ACCESSORIES</p> <p>(ADDITIONAL SPECIFICATIONS FOR <u>ITEM 24 – SURGICAL SET</u>)</p> <p>24.E – INFECTION CONTROL SUPPLIES: -ANTISEPTIC SOLUTIONS (E.G CHLORHEXIDINE OR IODINE-BASED SOLUTIONS) -ALCOHOL SWABS OR WIPES -STERILE GLOVES IN VARIOUS SIZES STERILE DISPOSABLE SURGICAL MASKS</p> |
| | <p>Section VII. TECHNICAL SPECIFICATIONS Page 25, Specification AMBULANCE EQUIPMENT & ACCESSORIES</p> <p>(ADDITIONAL SPECIFICATIONS FOR <u>ITEM 24 – SURGICAL SET</u>)</p> <p>24.F – WOUND CLOSURE MATERIALS: -SUTURES (E.G ABSORBABLE OR NON-ABSORBABLE SUTURES IN VARIOUS SIZES) -STERILE SUTURE NEEDLES -SKIN ADHESIVE (E.G TISSUE ADHESIVE OR MEDICAL GLUE) -STERILE WOUND CLOSURE STRIPS (E.G STERI-STRIPS)</p> |
| | <p>Section VII. TECHNICAL SPECIFICATIONS Page 25, Specification AMBULANCE EQUIPMENT & ACCESSORIES</p> <p>(ADDITIONAL SPECIFICATIONS FOR <u>ITEM 24 – SURGICAL SET</u>)</p> <p>24.G – BASIC SURGICAL EQUIPMENT: -SURGICAL DRAPES OR DISPOSABLE SURGICAL TRAYS -SURGICAL SCISSORS FOR CUTTING DRAPES OR SUTURES -SURGICAL MARKER FOR SURGICAL SITE IDENTIFICATION</p> |
| | <p>Section VII. TECHNICAL SPECIFICATIONS Page 25, Specification AMBULANCE EQUIPMENT & ACCESSORIES</p> <p>(ADDITIONAL SPECIFICATIONS FOR <u>ITEM 24 – SURGICAL SET</u>)</p> <p>24.H – PORTABLE LIGHTING:</p> |



Republic of the Philippines
Province of South Cotabato
City of Koronadal

Bids and Awards Committee

E-mail Address: lgubackor@gmail.com Website: www.philgeps.gov.ph and www.koronadal.gov.ph

| | |
|--|--|
| | <p>-PENLIGHTS OR HEADLAMPS</p> |
| | <p>Section VII. TECHNICAL SPECIFICATIONS Page 25, Specification AMBULANCE EQUIPMENT & ACCESSORIES</p> <p><u>(ADDITIONAL SPECIFICATIONS FOR ITEM 25 – EMERGENCY LIGHT (DOOR-MOUNTED))</u></p> <p>25.A – MOUNTING LOCATION: THE EMERGENCY LIGHT IS TYPICALLY MOUNTED ON THE EXTERIOR OF THE AMBULANCE, USUALLY ON THE ROOF OR THE UPPER PORTION OF THE SIDE DOOR(S).</p> |
| | <p>Section VII. TECHNICAL SPECIFICATIONS Page 25, Specification AMBULANCE EQUIPMENT & ACCESSORIES</p> <p><u>(ADDITIONAL SPECIFICATIONS FOR ITEM 25 – EMERGENCY LIGHT (DOOR-MOUNTED))</u></p> <p>25.B – LIGHTING CONFIGURATION: LED STROBE LIGHTS, ROTATING BEACONS, OR LED LIGHT BARS</p> |
| | <p>Section VII. TECHNICAL SPECIFICATIONS Page 25, Specification AMBULANCE EQUIPMENT & ACCESSORIES</p> <p><u>(ADDITIONAL SPECIFICATIONS FOR ITEM 25 – EMERGENCY LIGHT (DOOR-MOUNTED))</u></p> <p>25.C – COLORS & PATTERNS: RED & GREEN</p> |
| | <p>Section VII. TECHNICAL SPECIFICATIONS Page 25, Specification AMBULANCE EQUIPMENT & ACCESSORIES</p> <p><u>(ADDITIONAL SPECIFICATIONS FOR ITEM 25 – EMERGENCY LIGHT (DOOR-MOUNTED))</u></p> <p>25.D – POWER SOURCE: THE EMERGENCY LIGHT MAY BE POWERED BY THE AMBULANCE’S ELECTRICAL SYSTEM</p> |
| | <p>Section VII. TECHNICAL SPECIFICATIONS Page 25, Specification AMBULANCE EQUIPMENT & ACCESSORIES</p> <p><u>(ADDITIONAL SPECIFICATIONS FOR ITEM 26 – LIGHT BAR (RED, GREEN FLASHING LED))</u></p> |



Republic of the Philippines
Province of South Cotabato
City of Koronadal

Bids and Awards Committee

E-mail Address: lgubackor@gmail.com Website: www.philgeps.gov.ph and www.koronadal.gov.ph

| | |
|--|--|
| | <p><u>-COLOR: RED & GREEN</u> <u>-POWER: 12v / 24v</u> <u>-WITH BUILT-IN SIREN</u> <u>-WITH MIC MODULE & PUBLIC ADDRESS SYSTEM</u> <u>-INCLUDES BRACKET</u></p> |
| | <p>Section VII. TECHNICAL SPECIFICATIONS Page 25, Specification AMBULANCE EQUIPMENT & ACCESSORIES</p> <p><u>(ADDITIONAL SPECIFICATIONS FOR ITEM 27 – COMMUNICATION EQUIPMENT (INTERCOM SYSTEM))</u></p> <p>27.A CONTAINS TWO INTERCOMS</p> |
| | <p>Section VII. TECHNICAL SPECIFICATIONS Page 25, Specification AMBULANCE EQUIPMENT & ACCESSORIES</p> <p><u>(ADDITIONAL SPECIFICATIONS FOR ITEM 27 – COMMUNICATION EQUIPMENT (INTERCOM SYSTEM))</u></p> <p>27.B LED INDICATORS</p> |
| | <p>Section VII. TECHNICAL SPECIFICATIONS Page 25, Specification AMBULANCE EQUIPMENT & ACCESSORIES</p> <p><u>(ADDITIONAL SPECIFICATIONS FOR ITEM 27 – COMMUNICATION EQUIPMENT (INTERCOM SYSTEM))</u></p> <p>27.C VOLUME CONTROL</p> |
| | <p>Section VII. TECHNICAL SPECIFICATIONS Page 25, Specification AMBULANCE EQUIPMENT & ACCESSORIES</p> <p><u>(ADDITIONAL SPECIFICATIONS FOR ITEM 27 – COMMUNICATION EQUIPMENT (INTERCOM SYSTEM))</u></p> <p>27.D COMPACT SIZE</p> |
| | <p>Section VII. TECHNICAL SPECIFICATIONS Page 25, Specification AMBULANCE EQUIPMENT & ACCESSORIES</p> <p><u>(ADDITIONAL SPECIFICATIONS FOR ITEM 27 – COMMUNICATION EQUIPMENT (INTERCOM SYSTEM))</u></p> |



Republic of the Philippines
Province of South Cotabato
City of Koronadal

Bids and Awards Committee

E-mail Address: lgubackor@gmail.com Website: www.philgeps.gov.ph and www.koronadal.gov.ph

| | |
|--|--|
| | 27.E TWO KEYS: POWER & CALL (CALL/TALK) |
| | Section VII. TECHNICAL SPECIFICATIONS Page 25, Specification AMBULANCE EQUIPMENT & ACCESSORIES (ADDITIONAL SPECIFICATIONS FOR <u>ITEM 28 – 220 VOLT AC SUPPLY</u>) 28.A VOLTAGE: 220 VOLTS |
| | Section VII. TECHNICAL SPECIFICATIONS Page 25, Specification AMBULANCE EQUIPMENT & ACCESSORIES (ADDITIONAL SPECIFICATIONS FOR <u>ITEM 28 – 220 VOLT AC SUPPLY</u>) 28.B FREQUENCY: 50Hz OR 60Hz |
| | Section VII. TECHNICAL SPECIFICATIONS Page 25, Specification AMBULANCE EQUIPMENT & ACCESSORIES (ADDITIONAL SPECIFICATIONS FOR <u>ITEM 28 – 220 VOLT AC SUPPLY</u>) 28.C SAFETY FEATURES: PARAMOUNT, OFTEN INCLUDE VARIOUS SAFETY FEATURES. THESE MAY INCLUDE OVERLOAD PROTECTION, SHORT-CIRCUIT PROTECTION, GROUND FAULT PROTECTION & INSULATION TO PREVENT ELECTRIC SHOCK |
| | Section VII. TECHNICAL SPECIFICATIONS Page 25, Specification AMBULANCE EQUIPMENT & ACCESSORIES (ADDITIONAL SPECIFICATIONS FOR <u>ITEM 28 – 220 VOLT AC SUPPLY</u>) 28.D INVERTER OR GENERATOR: DEPENDING ON THE DESIGN OF THE AMBULANCES' ELECTRICAL SYSTEM, THE 220 VOLT AC SUPPLY MAY BE PROVIDED BY AN INVERTER CONVERTING DC POWER FROM BATTERIES OR BY A GENERATOR RUNNING ON FUEL SUCH AS DIESEL OR GASOLINE. |



Republic of the Philippines
Province of South Cotabato
City of Koronadal

Bids and Awards Committee

E-mail Address: lgubackor@gmail.com Website: www.philgeps.gov.ph and www.koronadal.gov.ph

| | |
|--|---|
| | <p>Section VII. TECHNICAL SPECIFICATIONS Page 25, Specification AMBULANCE EQUIPMENT & ACCESSORIES</p> <p>(ADDITIONAL SPECIFICATIONS FOR <u>ITEM 29 – 3 SITTER BENCH</u>: -MOUNTED ON AMBULANCE FLOOR: ONE ON THE HEAD SIDE OF THE PATIENT & ONE ON THE RIGHT SIDE OF THE PATIENT COMPARTMENT -MATERIALS: ALUMINUM FRAME AND PANELS COATED W/ WHITE COLOR PAINT OR OTHER LIGHT WEIGHT; THE SEAT & BACKSEAT MUST BE FITTED W/ MATTRESS MADE OF POLYURETHANE FOAM-COVERED W/ A WASHABLE LEATHERETTE MATERIAL. 2 POINT SEATBELT FOR EACH PERSON</p> |
| | <p>Section VII. TECHNICAL SPECIFICATIONS Page 25, Specification AMBULANCE EQUIPMENT & ACCESSORIES</p> <p>(ADDITIONAL SPECIFICATIONS FOR <u>ITEM 30 – PATIENT HANDLE BAR</u>: -STAINLESS STEEL</p> |

All other provisions of the bidding documents which are not affected shall remain in effect.

For guidance and information of all concerned.

SGD

Ar. En.P. MARVIN V. ALEGRIA, FUAP, PIEP
Chairperson – Bids & Awards Committee